

```
SmartSelect = "yes";
Coil 1 (exclude) = "None";
Uniformity = "CLEAR";
FOV RL (mm) = 256;
AP (mm) = 256;
FH (mm) = 160;
ACQ voxel size RL (mm) = 2;
AP (mm) = 2;
Slice thickness (mm) = 2;
Recon voxel size RL (mm) = 2;
AP (mm) = 2;
Fold-over suppression = "no";
Reconstruction matrix = 128;
SENSE = "yes";
P reduction (AP) = 2;
fMRI = "no";
Z Shim = "no";
k-t BLAST = "no";
Stacks = 1;
type = "parallel";
slices = 80;
slice gap = "user defined";
gap (mm) = 0;
slice orientation = "transverse";
fold-over direction = "AP";
fat shift direction = "P";
Stack Offc. AP (P=+mm) = -7.03728294;
RL (L=+mm) = -2.99954987;
FH (H=+mm) = 10.358182;
Ang. AP (deg) = 0;
RL (deg) = -0;
FH (deg) = -0;
Free rotatable = "no";
Minimum number of packages = 1;
Slice scan order = "default";
Large table movement = "no";
PlanAlign = "no";
REST slabs = 0;
Interactive positioning = "no";
External control = "no";
Patient position = "head first";
Patient body position = "head first";
Patient orientation = "supine";
Patient body orientation = "supine";
Scan type = "Imaging";
Scan mode = "MS";
technique = "SE";
Modified SE = "no";
Acquisition mode = "cartesian";
Fast Imaging mode = "EPI";
```

```

    shot mode = "single-shot";
Echoes = 1;
    partial echo = "no";
TE = "shortest";
Flip angle (deg) = 90;
TR = "shortest";
Halfscan = "yes";
    factor = 0.682539701;
Water-fat shift = "minimum";
Shim = "auto";
mDIXON = "no";
Fat suppression = "SPIR";
    strength = "strong";
    frequency offset = "user defined";
        offset (Hz) = 175;
Grad Rev Fat suppr = "no";
Water suppression = "no";
MTC = "no";
Research prepulse = "no";
Diffusion mode = "DTI";
    sequence = "SE";
    gradient duration = "maximum";
    gradient overplus = "no";
    directional resolution = "from file";
    average high b = "user defined";
        b-factor averages = "(0) 5",
            "(1000) 1", "", "", "", "",
            "", "", "", "", "", "", "",
            "", "", "", "", "", "", "",
            "", "", "", "", "", "", "",
            "", "", "", "", "", "", "",
            "", "", "", "", "", "";
Multi-transmit = "no";
SAR mode = "high";
B1 mode = "default";
SAR allow first level = "yes";
PNS mode = "moderate";
Gradient mode = "maximum";
SoftTone mode = "no";
Cardiac synchronization = "no";
Heart rate > 250 bpm = "no";
Respiratory compensation = "no";
Navigator respiratory comp = "no";
Flow compensation = "no";
Temporal slice spacing = "default";
NSA = 1;
Manual start = "yes";
Dynamic study = "no";
    dyn stabilization = "no";
Arterial Spin labeling = "no";
Preparation phases = "auto";

```

```

Interactive F0 = "no";
B0 field map = "no";
B1 field map = "no";
MIP/MPR = "no";
Images = "M", (3) "no";
Autoview image = "M";
Calculated images = (4) "no";
Reference tissue = "White matter";
Recon compression = "No";
Preset window contrast = "soft";
Reconstruction mode = "immediate";
Save raw data = "no";
Hardcopy protocol = "no";
Image filter = "system default";
Uniformity correction = "no";
Geometry correction = "default";
IF_info_seperator = 1634755923;
Total scan duration = "07:14.5";
Rel. SNR = 0.159877479;
Act. TR (ms) = "9245";
Act. TE (ms) = "76";
ACQ matrix M x P = "128 x 126";
ACQ voxel MPS (mm) = "2.00 / 2.03 / 2.00";
REC voxel MPS (mm) = "2.00 / 2.00 / 2.00";
Scan percentage (%) = 98.4375;
Packages = 1;
fMRI = "no";
Min. slice gap (mm) = -0;
Diffusion gradient timing DELTA / delta (ms) =
"37.0 / 22.2";
User defined DTI scheme = "Rush (41, 1000)";
EPI factor = 63;
WFS (pix) / BW (Hz) = "20.231 / 21.5";
BW in EPI freq. dir. (Hz) = "2095.5";
SPIR offset act./default (Hz) = "175 [220]";
Head SAR = "< 32 %";
Whole body SAR / level = "< 0.1 W/kg / normal";
SED = " 0.0 kJ/kg";
Max B1+rms = "1.33 uT";
PNS / level = "79 % / normal";
dB/dt = "67.6 T/s";
Sound Pressure Level (dB) = 19.776104;

```