

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\\GROSENBERG\\Ugrant_20294\\Ugrant_harmonization\\unm_cvr_09072018

TA: 7:17 PAT: 2 Voxel size: 3.4x3.4x3.8 mm Rel. SNR: 1.00 USER: ep2d_complex

Properties		Series	Ascending
Prio Recon	Off	Special sat.	None
Before measurement	System		
After measurement	On	Body	Off
Load to viewer	Off	HEP	On
Inline movie	On	HEA	On
Auto store images	Off	Positioning mode	FIX
Load to stamp segments	Off	Table position	H
Load images to graphic segments	Off	Table position	0 mm
Auto open inline display	Off	MSMA	S - C - T
Start measurement without further preparation	On	Sagittal	R >> L
Wait for user to start	Off	Coronal	A >> P
Start measurements	single	Transversal	F >> H
Routine		Coil Combine Mode	Adaptive Combine
Slice group 1		Auto Coil Select	Default
Slices	34	Shim mode	Standard
Dist. factor	0 %	Adjust with body coil	Off
Position	R2.3 A20.9 F29.0	Confirm freq. adjustment	Off
Orientation	Transversal	Assume Silicone	Off
Phase enc. dir.	A >> P	? Ref. amplitude 1H	0.000 V
Rotation	0.00 deg	Adjustment Tolerance	Auto
Phase oversampling	0 %	Adjust volume	
FoV read	220 mm	Position	R2.3 A20.9 F29.0
FoV phase	100.0 %	Orientation	Transversal
Slice thickness	3.8 mm	Rotation	0.00 deg
TR	1500 ms	R >> L	220 mm
TE	21 ms	A >> P	220 mm
Averages	1	F >> H	130 mm
Concatenations	1		
Filter	Prescan Normalize	Physio	
Coil elements	HEA;HEP	1st Signal/Mode	None
Contrast		BOLD	
MTC	Off	GLM Statistics	Off
Flip angle	90 deg	Dynamic t-maps	Off
Fat suppr.	Fat sat.	Starting ignore meas	0
Averaging mode	Long term	Ignore after transition	0
Reconstruction	Magnitude	Model transition states	Off
Measurements	287	Temp. highpass filter	Off
Delay in TR	0 ms	Threshold	4.00
Multiple series	Off	Paradigm size	20
Resolution		Meas[1]	Baseline
Base resolution	64	Meas[2]	Baseline
Phase resolution	100 %	Meas[3]	Baseline
Phase partial Fourier	Off	Meas[4]	Baseline
Interpolation	Off	Meas[5]	Baseline
PAT mode	GRAPPA	Meas[6]	Baseline
Accel. factor PE	2	Meas[7]	Baseline
Ref. lines PE	24	Meas[8]	Baseline
Matrix Coil Mode	Auto (Triple)	Meas[9]	Baseline
Reference scan mode	Separate	Meas[10]	Baseline
Distortion Corr.	Off	Meas[11]	Active
Unfiltered images	Off	Meas[12]	Active
Prescan Normalize	On	Meas[13]	Active
Raw filter	On	Meas[14]	Active
Elliptical filter	Off	Meas[15]	Active
Hamming	Off	Meas[16]	Active
Geometry		Meas[17]	Active
Multi-slice mode	Interleaved	Meas[18]	Active
		Meas[19]	Active
		Meas[20]	Active
		Motion correction	Off
		Spatial filter	Off

SIEMENS MAGNETOM TrioTim syngo MR B17

Sequence

Introduction	Off
Bandwidth	2442 Hz/Px
Free echo spacing	Off
Echo spacing	0.49 ms
EPI factor	64
RF pulse type	Normal
Gradient mode	Fast