

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\ACAPRIHAN\phantom_20200\ugrant_adni_phantom\unm_mprage_adni_12ch_qc

TA: 6:03 PAT: 2 Voxel size: 1.0x1.0x1.2 mm Rel. SNR: 1.00 USER: tfl_mgh_multiecho

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	L1.3 P0.0 H24.3
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	176
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.20 mm
TR	2530 ms
TE 1	1.64 ms
TE 2	3.5 ms
TE 3	5.36 ms
TE 4	7.22 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

Magn. preparation	Non-sel. IR
T1	1200 ms
Flip angle	7.0 deg
Fat suppr.	None
Water suppr.	None
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	256
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32
Accel. factor 3D	1
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Integrated

Image Filter	Off
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Single shot
Series	Interleaved

System

Body	Off
HEP	On
HEA	On

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
Auto Coil Select	Default

Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Dark blood	Off

Inline

Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	3D
Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	4
Bandwidth 1	651 Hz/Px

SIEMENS MAGNETOM TrioTim syngo MR B17

Bandwidth 2	651 Hz/Px
Bandwidth 3	651 Hz/Px
Bandwidth 4	651 Hz/Px
Flow comp. 1	No
Flow comp. 2	No
Flow comp. 3	No
Flow comp. 4	No
Echo spacing	10.3 ms

RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On

Readout polarity	Positive
Readout trajectory	Bipolar
Add. scale factor	4.0
Gradient spoiling	Integral
Gradient moment factor	3.0
Siemens reconstruction	On
Save raw k-space data	Off
Averaging	RMS